

Equilibrium Relative Humidity Saturated Salt Solutions

Relative Humidity (%RH)			
Temperature °C	Lithium Chloride	Potassium Acetate	Magnesium Chloride
0	11.23 ± 0.54	23.28 ± 0.53 23.40 ± 0.32 23.11 ± 0.25 22.51 ± 0.32 21.61 ± 0.53	33.66 ± 0.33
5	11.26 ± 0.47		33.60 ± 0.28
10	11.29 ± 0.41		33.47 ± 0.24
15	11.30 ± 0.35		33.30 ± 0.21
20	11.31 ± 0.31		33.07 ± 0.18
25	11.30 ± 0.27		32.78 ± 0.16
30	11.28 ± 0.24		32.44 ± 0.14
35	11.25 ± 0.22		32.05 ± 0.13
40	11.21 ± 0.21		31.60 ± 0.13
45	11.16 ± 0.21		31.10 ± 0.13
50	11.10 ± 0.22	29.93 ± 0.16 29.26 ± 0.18 28.54 ± 0.21 27.77 ± 0.25 26.94 ± 0.29 26.05 ± 0.34 25.11 ± 0.39 24.12 ± 0.46 23.07 ± 0.52 21.97 ± 0.60	30.54 ± 0.13
55	11.03 ± 0.23		29.93 ± 0.16
60	10.95 ± 0.26		29.26 ± 0.18
65	10.86 ± 0.29		28.54 ± 0.21
70	10.75 ± 0.33		27.77 ± 0.25
75	10.64 ± 0.38		26.94 ± 0.29
80	10.51 ± 0.44		26.05 ± 0.34
85	10.38 ± 0.51		25.11 ± 0.39
90	10.23 ± 0.59		24.12 ± 0.46
95	10.07 ± 0.67		23.07 ± 0.52
100	9.90 ± 0.77		21.97 ± 0.60

Saturated Salt Solutions

A very convenient method to calibrate humidity sensors is the use of saturated salt solutions. At any temperature, the concentration of a saturated solution is fixed and does not have to be determined. By providing excess solute, the solution will remain saturated even in the presence of modest moisture sources and sinks. When the solute is a solid in the pure phase, it is easy to determine that there is saturation.

The saturated salt solution, made up as a slushy mixture with distilled water and chemically pure salt, is enclosed in a sealed metal or a glass chamber. Wexler and Hasegawa measured the humidity in the atmosphere above eight saturated salt solutions for ambient temperatures 0 to 50°C using a dew-point hygrometer. Later, Greenspan compiled, from the literature, data on 28 saturated salt solutions to cover the entire range of relative humidity. Using a data base from 21 separate investigations comprising 1106 individual measurements, fits were made by the method of least squares to regular polynomial equations to obtain the "best" value of relative humidity in air as a function of temperature. These values are summarized in the table shown. 

Relative Humidity (%RH)						
Temperature °C	Potassium Carbonate	Magnesium Nitrate	Sodium Chloride	Potassium Chloride	Potassium Nitrate	Potassium Sulfate
0	43.13 ± 0.66	60.35 ± 0.55	75.51 ± 0.34	88.61 ± 0.53	96.33 ± 2.9	98.77 ± 1.1
5	43.13 ± 0.50	58.86 ± 0.43	75.65 ± 0.27	87.67 ± 0.45	96.27 ± 2.1	98.48 ± 0.91
10	43.14 ± 0.39	57.36 ± 0.33	75.67 ± 0.22	86.77 ± 0.39	95.96 ± 1.4	98.18 ± 0.76
15	43.15 ± 0.33	55.87 ± 0.27	75.61 ± 0.18	85.92 ± 0.33	95.41 ± 0.96	97.89 ± 0.63
20	43.16 ± 0.33	54.38 ± 0.23	75.47 ± 0.14	85.11 ± 0.29	94.62 ± 0.66	97.59 ± 0.53
25	43.16 ± 0.39	52.89 ± 0.22	75.29 ± 0.12	84.34 ± 0.26	93.58 ± 0.55	97.30 ± 0.45
30	43.17 ± 0.50	51.40 ± 0.24	75.09 ± 0.11	83.62 ± 0.25	92.31 ± 0.60	97.00 ± 0.40
35		49.91 ± 0.29	74.87 ± 0.12	82.95 ± 0.25	90.79 ± 0.83	96.71 ± 0.38
40		48.42 ± 0.37	74.68 ± 0.13	82.32 ± 0.25	89.03 ± 1.2	96.41 ± 0.38
45		46.93 ± 0.47	74.52 ± 0.16	81.74 ± 0.28	87.03 ± 1.8	96.12 ± 0.40
50		45.44 ± 0.60	74.43 ± 0.19	81.20 ± 0.31	84.78 ± 2.5	95.82 ± 0.45
55			74.41 ± 0.24	80.70 ± 0.35		
60			74.50 ± 0.30	80.25 ± 0.41		
65			74.71 ± 0.37	79.85 ± 0.48		
70			75.06 ± 0.45	79.49 ± 0.57		
75			75.58 ± 0.55	79.17 ± 0.66		
80			76.29 ± 0.65	78.90 ± 0.77		
85				78.68 ± 0.89		
90				78.50 ± 1.0		
95						
100						



UNITED STATES

www.omega.com

1-800-TC-OMEGA
Stamford, CT.

CANADA

www.omega.ca

Laval(Quebec)
1-800-TC-OMEGA

GERMANY

www.omega.de

Deckenpfronn, Germany
0800-8266342

UNITED KINGDOM

www.omega.co.uk

Manchester, England
0800-488-488

FRANCE

www.omega.fr

Guyancourt, France
088-466-342

CZECH REPUBLIC

www.omegaeng.cz

Karviná, Czech Republic
596-311-899

BENELUX

www.omega.nl

Amstelveen, NL
0800-099-33-44



More than 100,000 Products Available!

• Temperature

Calibrators, Connectors, General Test and Measurement Instruments, Glass Bulb Thermometers, Handheld Instruments for Temperature Measurement, Ice Point References, Indicating Labels, Crayons, Cements and Lacquers, Infrared Temperature Measurement Instruments, Recorders Relative Humidity Measurement Instruments, RTD Probes, Elements and Assemblies, Temperature & Process Meters, Timers and Counters, Temperature and Process Controllers and Power Switching Devices, Thermistor Elements, Probes and Assemblies, Thermocouples Thermowells and Head and Well Assemblies, Transmitters, Wire

• Flow and Level

Air Velocity Indicators, Doppler Flowmeters, Level Measurement, Magnetic Flowmeters, Mass Flowmeters, Pitot Tubes, Pumps, Rotameters, Turbine and Paddle Wheel Flowmeters, Ultrasonic Flowmeters, Valves, Variable Area Flowmeters, Vortex Shedding Flowmeters

• pH and Conductivity

Conductivity Instrumentation, Dissolved Oxygen Instrumentation, Environmental Instrumentation, pH Electrodes and Instruments, Water and Soil Analysis Instrumentation

• Data Acquisition

Auto-Dialers and Alarm Monitoring Systems, Communication Products and Converters, Data Acquisition and Analysis Software, Data Loggers Plug-in Cards, Signal Conditioners, USB, RS232, RS485 and Parallel Port Data Acquisition Systems, Wireless Transmitters and Receivers

• Pressure, Strain and Force

Displacement Transducers, Dynamic Measurement Force Sensors, Instrumentation for Pressure and Strain Measurements, Load Cells, Pressure Gauges, Pressure Reference Section, Pressure Switches, Pressure Transducers, Proximity Transducers, Regulators, Strain Gages, Torque Transducers, Valves

• Heaters

Band Heaters, Cartridge Heaters, Circulation Heaters, Comfort Heaters, Controllers, Meters and Switching Devices, Flexible Heaters, General Test and Measurement Instruments, Heater Hook-up Wire, Heating Cable Systems, Immersion Heaters, Process Air and Duct, Heaters, Radiant Heaters, Strip Heaters, Tubular Heaters